



Company Name	BONSAI JACK LLC	Calibration Date	11-9-18
Address	17940 BALFOUR TERRACE	Next Calibration Due Date	NEW
City/State/Zip	FT MYERS, FL 33913	Service Report #	1154W

THE WEIGHING SYSTEM LISTED BELOW HAS BEEN CALIBRATED USING TEST STANDARDS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST)

EQUIPMENT INFORMATION

Scale Type	BENCH	Manufacturer:	OHAUS	Max Capacity:	3000 G
Customer ID#	786-262-3135	Model:	V71P3T	Min Division:	1 G
Location		Serial #:	8339360245	Units / Class:	III

TEST RESULTS

AS FOUND:

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
2000	2000	0
3000	3000	0

AS LEFT:

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
2000	2000	0
3000	3000	0

TOLERANCE:

Tolerance Limit	
Low	High
0	0
0	2
99	101
998	1002
1998	2002
2997	3003


CORNER TEST:

Corner 1	1000
Corner 2	1000
Corner 3	1000
Corner 4	1000

Indicator

1	2
4	3

Pass Fail



Yes No

CLASS III Maintenance TOLERANCE	
Acceptance Tolerance = 1/2 of Maintenance Tolerance	
0-500 Division Range	= 1 Divisions Allowable Error
501-2000 Division Range	= 2 Divisions Allowable Error
2001-4000 Division Range	= 3 Divisions Allowable Error
4001-5000 Division Range	= 5 Divisions Allowable Error
Max Division Range = Capacity / Min Division	
Example: 100 / .02 = 5000 Max Divisions	
Tolerance Limit = Test Load / Min Division	
Example: 25 / .02 = 1250 Division Range (Allowable Error = 2 divisions or +/- .04)	

REPEATABILITY TEST

Test Weight	Reading 1	Zero	Reading 2	Zero	Reading 3	Zero
1000	1000	0	1000	0	1000	0

Comments / Observations:

SCALE IS WEIGHING WITHIN TOLERANCE

TEST STANDARDS USED: (Weight or Kit Serial Numbers)

Certification Performed By: WAYNE M.