



Company Name	BONSAI JACK LLC	Calibration Date	11-9-18
Address	17940 BALFOUR TERRACE	Next Calibration Due Date	NEW
City/State/Zip	FORT MYERS, FL 33913	Service Report #	1154W

THE WEIGHING SYSTEM LISTED BELOW HAS BEEN CALIBRATED USING TEST STANDARDS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST)

**EQUIPMENT INFORMATION**

Scale Type	BENCH	Manufacturer:	OHAUS	Max Capacity:	6000 G
Customer ID#	786262-3135	Model:	V71P6T	Min Division:	2 G
Location		Serial #:	8339350377	Units / Class:	III

**TEST RESULTS**

**AS FOUND:**

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
3000	3000	0
6000	6000	0

**AS LEFT:**

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
3000	3000	0
6000	6000	0

**TOLERANCE:**

Tolerance Limit	
Low	High
0	0
0	0
98	102
998	1002
2998	3004
5994	6006


**CORNER TEST:**

Corner 1	2000
Corner 2	2000
Corner 3	2000
Corner 4	2000

**Indicator**

1	2
4	3

Pass  Fail



Yes  No

CLASS III Maintenance TOLERANCE	
Acceptance Tolerance = 1/2 of Maintenance Tolerance	
0-500 Division Range	= 1 Divisions Allowable Error
501-2000 Division Range	= 2 Divisions Allowable Error
2001-4000 Division Range	= 3 Divisions Allowable Error
4001-5000 Division Range	= 5 Divisions Allowable Error
Max Division Range = Capacity / Min Division	
Example: 100 / .02 = 5000 Max Divisions	
Tolerance Limit = Test Load / Min Division	
Example: 25 / .02 = 1250 Division Range (Allowable Error = 2 divisions or + .04)	

**REPEATABILITY TEST**

Test Weight	Reading 1	Zero	Reading 2	Zero	Reading 3	Zero
2000	2000	0	2000	0	2000	0

**Comments / Observations:**

SCALE IS WEIGHING WITHIN TOLERANCE

**TEST STANDARDS USED: (Weight or Kit Serial Numbers)**

Certification Performed By: WAYNE M.