



5740 Zip Drive Fort Myers, Florida 33905 239-219-6803



Calibration Certificate

Florida Dept. of Agriculture Registration # 490

Company Name	BONSAI JACK LLC	Calibration Date	2-14-20
Address	17940 BALFOUR TERRACE	Next Calibration Due Date	T&C
City/State/Zip	FT MYERS, FL 33913	Service Report #	1612W

THE WEIGHING SYSTEM LISTED BELOW HAS BEEN CALIBRATED USING TEST STANDARDS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST)

EQUIPMENT INFORMATION

Scale Type	BENCH	Manufacturer:	OHAUS	Max Capacity:	5000 g
Customer ID#	786-262-3135	Model:	V71P3T	Min Division:	1 g
Location		Serial #:	8330360245	Units / Class:	g

TEST RESULTS

AS FOUND:

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
2000	2000	0
3000	3000	0

AS LEFT:

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
2000	2000	0
3000	3000	0

TOLERANCE:

Tolerance Limit	
Low	High
0	0
0	0
96	100
998	1000
1998	2000
2997	3000

CORNER TEST:

Corner 1	1000
Corner 2	1000
Corner 3	1000
Corner 4	1000

Indicator

1	2
4	3

CLASS III Maintenance TOLERANCE

Acceptance Tolerance = % of Maintenance Tolerance

0-500 Division Range = 1 Divisions Allowable Error
 501-2000 Division Range = 2 Divisions Allowable Error
 2001-4000 Division Range = 3 Divisions Allowable Error
 4001-5000 Division Range = 5 Divisions Allowable Error

Max Division Range = Capacity / Min Division
 Example: 100 / .02 = 5000 Max Divisions

Tolerance Limit = Test Load / Min Division
 Example: 25 / .02 = 1250 Division Range (Allowable Error = 2 divisions or + .04)

Pass Fail

Yes No

REPEATABILITY TEST

Test Weight	Reading 1	Zero	Reading 2	Zero	Reading 3	Zero
1000	1000	0	1000	0	1000	0

Comments / Observations:

SCALE IS WEIGHING WITHIN TOLERANCE

TEST STANDARDS USED: (Weight or Kit Serial Numbers)

Certification Performed By: WAYNE M.