



Company Name	BONSAI JACK LLC	Calibration Date	06 02 2021
Address	17940 BALFOUR TERRACE	Next Calibration Due Date	06 02 2022
City/State/Zip	FT MYERS, FL 33913	Service Report #	

THE WEIGHING SYSTEM LISTED BELOW HAS BEEN CALIBRATED USING TEST STANDARDS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST)

**EQUIPMENT INFORMATION**

Scale Type	BENCH	Manufacturer	OHAUS	Max Capacity	3000 G
Customer ID#	786-262-3135	Model	V71P3T	Min Division	1
Location		Serial #	8339360245	Units / Class	grams I

**TEST RESULTS**

**AS FOUND:**

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
2000	2000	0
3000	3000	0

**AS LEFT:**

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
2000	2000	0
3000	3000	0

**TOLERANCE:**

Tolerance Limit	
Low	High
0	0
1	1
100	100
1000	1000
2000	2000
3000	3000


**CORNER TEST:**

Corner 1	1000.0
Corner 2	1000.0
Corner 3	1000.0
Corner 4	1000.0

**Indicator**

1	2
4	3

Pass  Fail



Yes  No

CLASS III Maintenance TOLERANCE	
Acceptance Tolerance = 1/2 of Maintenance Tolerance	
0-500 Division Range	= 1 Divisions Allowable Error
501-2000 Division Range	= 2 Divisions Allowable Error
2001-4000 Division Range	= 3 Divisions Allowable Error
4001-5000 Division Range	= 5 Divisions Allowable Error
Max Division Range = Capacity / Min Division	
Example: 100 / .02 = 5000 Max Divisions	
Tolerance Limit = Test Load / Min Division	
Example: 25 / .02 = 1250 Division Range (Allowable Error = 2 divisions or +/- .04)	

**REPEATABILITY TEST**

Test Weight	Reading 1	Zero	Reading 2	Zero	Reading 3	Zero
1000	1000.0	0	1000.0	0	1000.0	0

Comments / Observations:

SCALE IS WEIGHING WITHIN TOLERANCE

**TEST STANDARDS USED: (Weight or Kit Serial Numbers)**

TS 1327

Certification Performed By: Daniel/Zach