



Company Name	BONSAI JACK LLC	Calibration Date	06 02 2021
Address	17940 BALFOUR TERRACE	Next Calibration Due Date	06 02 2022
City/State/Zip	FT MYERS, FL 33913	Service Report #	

THE WEIGHING SYSTEM LISTED BELOW HAS BEEN CALIBRATED USING TEST STANDARDS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST)

EQUIPMENT INFORMATION

Scale Type	BENCH	Manufacturer:	OHAUS	Max Capacity:	3000 G
Customer ID#	786-262-3135	Model:	V71P3T	Min Division:	1
Location		Serial #:	8339350377	Units / Class:	grams I

TEST RESULTS

AS FOUND:

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
2000	2000	0
3000	3000	0

AS LEFT:

Test Load	Scale Indication	Error
0	0	0
1	1	0
100	100	0
1000	1000	0
2000	2000	0
3000	3000	0

TOLERANCE:

Tolerance Limit	
Low	High
0	0
0	0
96	96
996	996
1996	1996
2997	2997


CORNER TEST:

Corner 1	1000.0
Corner 2	1000.0
Corner 3	1000.0
Corner 4	1000.0

Indicator

1	2
4	3

Pass Fail



Yes No

CLASS III Maintenance TOLERANCE

Acceptance Tolerance = 1/2 of Maintenance Tolerance

0-500 Division Range = 1 Divisions Allowable Error

501-2000 Division Range = 2 Divisions Allowable Error

2001-4000 Division Range = 3 Divisions Allowable Error

4001-5000 Division Range = 5 Divisions Allowable Error

Max Division Range = Capacity / Min Division

Example: 100 / .02 = 5000 Max Divisions

Tolerance Limit = Test Load / Min Division

Example: 25 / .02 = 1250 Division Range (Allowable Error = 2 divisions or +/- .04)

REPEATABILITY TEST

Test Weight	Reading 1	Zero	Reading 2	Zero	Reading 3	Zero
1000	1000.0	0	1000.0	0	1000.0	0

Comments / Observations:

SCALE IS WEIGHING WITHIN TOLERANCE

TEST STANDARDS USED: (Weight or Kit Serial Numbers)

TS 1327

Certification Performed By:

Daniel/Zach