



Company Name	Bonzai Jack LLC	Calibration Date	2-3-22
Address	1794 Balfour Terrace	Next Calibration Due Date	2-3-23
City/State/Zip	Fort Myers, FL 33913	Service Report #	

THE WEIGHING SYSTEM LISTED BELOW HAS BEEN CALIBRATED USING TEST STANDARDS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST)

EQUIPMENT INFORMATION

Scale Type	bench	Manufacturer:	Ohaus	Max Capacity:	6
Customer ID#		Model:	Valor 7000	Min Division:	.002
Location		Serial #:	8339360245	Units / Class:	Lbs I

TEST RESULTS

AS FOUND:

Test Load	Scale Indication	Error
0	0	0
1	1	0
2	2	0
3	3	0
5	5	0

AS LEFT:

Test Load	Scale Indication	Error
Same		

TOLERANCE:

Tolerance Limit	
Low	High


CORNER TEST:

Corner 1	5
Corner 2	5
Corner 3	5
Corner 4	5

Indicator

1	2
4	3

Pass xxx Fail



Yes xxx No

CLASS III Maintenance TOLERANCE	
Acceptance Tolerance = 1/2 of Maintenance Tolerance	
0-500 Division Range	= 1 Divisions Allowable Error
501-2000 Division Range	= 2 Divisions Allowable Error
2001-4000 Division Range	= 3 Divisions Allowable Error
4001-5000 Division Range	= 5 Divisions Allowable Error
Max Division Range = Capacity / Min Division	
Example: 100 / .02 = 5000 Max Divisions	
Tolerance Limit = Test Load / Min Division	
Example: 25 / .02 = 1250 Division Range (Allowable Error = 2 divisions or + -.04)	

REPEATABILITY TEST

Test Weight	Reading 1	Zero	Reading 2	Zero	Reading 3	Zero
10	10	0	10	0	10	0

Comments / Observations:

Scale is in tolerance at this time

TEST STANDARDS USED: (Weight or Kit Serial Numbers)

Certification Performed By: Zach/David