



2130 Park 82 Drive, Fort Myers, Florida 33905 239-334-1717



# Calibration Certificate

Florida Dept. of Agriculture Registration # 490

Company Name	Bonzai Jack LLC	Calibration Date	2-3-22
Address	1794 Balfour Terrace	Next Calibration Due Date	2-3-23
City/State/Zip	Fort Myers, FL 33913	Service Report #	

THE WEIGHING SYSTEM LISTED BELOW HAS BEEN CALIBRATED USING TEST STANDARDS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST)

## EQUIPMENT INFORMATION

Scale Type	bench	Manufacturer:	Ohaus	Max Capacity:	15
Customer ID#		Model:	Valor 7300	Min Division:	0005
Location		Serial #:	8839350377	Units / Class:	Lbs

## TEST RESULTS

### AS FOUND:

Test Load	Scale Indication	Error
0	0	0
1	1	0
5	5	0
10	10	0
15	15	0

### AS LEFT:

Test Load	Scale Indication	Error
Same		

### TOLERANCE:

Tolerance Limit	
Low	High


### CORNER TEST:

Corner 1	5
Corner 2	5
Corner 3	5
Corner 4	5

### Indicator

1	2
4	3

Pass  Fail



Yes  No

CLASS III Maintenance TOLERANCE	
Acceptance Tolerance = 1/2 of Maintenance Tolerance	
0-500 Division Range	= 1 Divisions Allowable Error
501-2000 Division Range	= 2 Divisions Allowable Error
2001-4000 Division Range	= 3 Divisions Allowable Error
4001-5000 Division Range	= 5 Divisions Allowable Error
Max Division Range = Capacity / Min Division	
Example: 100 / .02 = 5000 Max Divisions	
Tolerance Limit = Test Load / Min Division	
Example: 25 / .02 = 1250 Division Range (Allowable Error = 2 divisions or +-.04)	

### REPEATABILITY TEST

Test Weight	Reading 1	Zero	Reading 2	Zero	Reading 3	Zero
10	10	0	10	0	10	0

### Comments / Observations:

Scale is in tolerance at this time.

### TEST STANDARDS USED: (Weight or Kit Serial Numbers)


Certification Performed By: Zach/David